



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/669,883	09/23/2003	Scott Manalis	0492611-0510 (MIT 10443)	2224
24280	7590	09/28/2006	EXAMINER	
CHOATE, HALL & STEWART LLP TWO INTERNATIONAL PLACE BOSTON, MA 02110			FORMAN, BETTY J	
			ART UNIT	PAPER NUMBER
			1634	

DATE MAILED: 09/28/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/669,883

Applicant(s)

MANALIS ET AL.

Examiner

BJ Forman

Art Unit

1634

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 27 July 2006.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-7,9-17,19-21,28,43,44,46 and 162-165 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-7,9-17,19-21,28,43,44,46 and 162-165 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☒ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☐ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☐ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date _____
- ☐ Notice of Informal Patent Application
- ☐ Other: _____

DETAILED ACTION

Continued Examination Under 37 CFR 1.114

1. A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on 27 July 2006 has been entered.

Status of the Claims

2. This action is in response to papers filed 27 July 2006 in which A Declaration under 37 C.F.R. § 1.132 was submitted, Terminal Disclaimers were submitted, claim 1 was amended and claims 7, 18, 22-27, 29-42 were canceled.

The amendments and Terminal Disclaimers have been thoroughly reviewed and entered. The previous rejections in the Office Action dated 21 March 2006, not reiterated below, are withdrawn in view of the amendments or Terminal Disclaimers. Applicant's arguments have been thoroughly reviewed and are discussed below as they apply to the instant grounds for rejection. New grounds for rejection are discussed.

Claims 1-7, 9-17, 19-21, 28, 43-44, 46 and 162-165 are under prosecution.

Declaration under 37 C.F.R. § 1.132

3. The declaration under 37 CFR 1.132 filed 27 July 2006 is insufficient to overcome the rejections of claims drawn to microfluidic channels based upon at least Boisseau and Pinkel, as set forth in the last Office action because: The declaration merely states the inventor own opinion as to the meaning, dimensions and volume encompassed by the terms microfluidics

Art Unit: 1634

and microfluidic channels i.e. "volumes....in the order of nanoliter or picoliters"; "height and/or depth.....between about a fraction of a micron and about 100 microns".

The declaration does not provide any evidence to support the inventor's opinion. As such, the declaration merely provides allegations of facts. Furthermore, the asserted volumes and dimensions within the nano and pico range seem to contradict the plain meaning of the claimed microfluidic channel.

Specification

Reiterated from previous Office Action

4. The amendment filed 22 July 2005 is objected to under 35 U.S.C. 132(a) because it introduces new matter into the disclosure. 35 U.S.C. 132(a) states that no amendment shall introduce new matter into the disclosure of the invention. The added material which is not supported by the original disclosure is as follows:

The amendment introduces a first paragraph cross-referencing co-pending application 10/336,549. The amendment states that this application is a "Continuation-in-Part of the co-pending application and claims priority to provisional application 60/405,184. The first paragraph and priority claim are deemed new matter. 37 C.F.R. § 1.78(ii)-(iii) states:

This reference must be submitted during the pendency of the later-filed application. If the later-filed application is an application filed under 35 U.S.C. 111(a), this reference must also be submitted within the later of four months from the actual filing date of the later-filed application or sixteen months from the filing date of the prior-filed application. If the later-filed application is a nonprovisional application which entered the national stage from an international application after compliance with 35 U.S.C. 371, this reference must also be submitted within the later of four months from the date on which the national stage commenced under 35 U.S.C. 371 (b) or (f) in the later-filed international application or sixteen months from the filing date of the prior-filed application. These time periods are not extendable. Except as provided in paragraph (a)(3) of this section, the failure to timely submit the reference required by 35 U.S.C. 120 and paragraph (a)(2)(i) of this section is considered a waiver of any benefit under 35 U.S.C. 120, 121, or 365(c) to such prior-filed application.

Art Unit: 1634

The reference was not timely submitted. Therefore, the amendment constitutes new matter.

Response to Applicant

5. Applicant states that a petition to claim priority was submitted 20 July 2005. The petition has been reviewed and dismissed 16 March 2006. Therefore the above rejection is maintained.

Applicant is required to cancel the new matter in the reply to this Office Action.

Claim Rejections - 35 USC § 112

6. The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

7. Claims 1-7, 9-17, 19-21, 28, 43-44, 46 and 162-165 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter which was not described in the specification in such a way as to reasonably convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention.

Claim 1 (from which all other claims depend) has been amended to define a beam that "encases" one or more sealed microfluidic channels. Applicant has not pointed to support for the newly defined beam and a review of the specification reveals no such support. The specification describes enclosing the beam to provide a protected environment (page 18, second full paragraph) and further describes solution enclosed within the microfluidic channels (page 24, second full paragraph). However, the specification does not teach or describe the newly claimed beam encasing one AND/OR multiple microfluidic channels.

Art Unit: 1634

MPEP 2163.06 notes "If NEW MATTER IS ADDED TO THE CLAIMS, THE EXAMINER SHOULD REJECT THE CLAIMS UNDER 35 U.S.C. 112, FIRST PARAGRAPH - WRITTEN DESCRIPTION REQUIREMENT. *IN RE RASMUSSEN*, 650 F.2d 1212, 211 USPQ 323 (CCPA 1981)." MPEP 2163.02 teaches that "Whenever the issue arises, the fundamental factual inquiry is whether a claim defines an invention that is clearly conveyed to those skilled in the art at the time the application was filed...If a claim is amended to include subject matter, limitations, or terminology not present in the application as filed, involving a departure from, addition to, or deletion from the disclosure of the application as filed, the examiner should conclude that the claimed subject matter is not described in that application." MPEP 2163.06 further notes "WHEN AN AMENDMENT IS FILED IN REPLY TO AN OBJECTION OR REJECTION BASED ON 35 U.S.C. 112, FIRST PARAGRAPH, A STUDY OF THE ENTIRE APPLICATION IS OFTEN NECESSARY TO DETERMINE WHETHER OR NOT "NEW MATTER" IS INVOLVED. APPLICANT SHOULD THEREFORE SPECIFICALLY POINT OUT THE SUPPORT FOR ANY AMENDMENTS MADE TO THE DISCLOSURE" (emphasis added).

Claim Rejections - 35 USC § 102

8. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

- (a) the invention was known or used by others in this country, or patented or described in a printed publication in this or a foreign country, before the invention thereof by the applicant for a patent.
- (b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.
- (c) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

9. Claims 1, 2, 162 and 165 are rejected under 35 U.S.C. 102(b) as being anticipated by Boisseau et al (U.S. Patent No. 4,896,966, issued 30 January 1990).

Regarding Claim 1, Boisseau et al disclose an apparatus for detecting an analyte comprising a suspended beam encasing a microfluidic channel (motility scanner #50) and connected to at least one mechanically stable support (support #16) wherein the channel has

Art Unit: 1634

at least one surface (e.g. diluent e.g. Hepes) that reacts with the analyte (e.g. suspend sperm) (Column 7, lines 40-50 and fig. 1-2).

Regarding Claim 2, Boisseau et al disclose the apparatus of Claim 1. The instant claim recites "the microfluidic channel is treated with a capture ligand that binds the analyte". This is a recitation of intended use for the apparatus but does not define or further limit the structure of the apparatus.

Regarding Claim 162, Boisseau et al disclose the apparatus wherein the beam is in a controlled environment i.e. portable housing #26 (Column 3, lines 35-40).

Regarding Claim 165, Boisseau et al disclose the apparatus wherein the beam is suspended between two mechanically stable supports (#110, Fig. 7).

Response to Arguments

10. Applicant argues that Biosseau et al do not teach a microfluidic channel because the dimension of their channel is roughly "orders of magnitude ($\sim 10^5$) larger than sub-nanoliter scale of microfluidics." Applicant points to a website microfluidics is defined as devices and processes dealing with fluids on the order of nanoliters and picoliters.

The argument has been considered but is not found persuasive because the claimed channels are not limited or defined by specific dimensions. The claims merely define the channel as microfluidic. Furthermore, any dimensions can be defined on the micro or sub-nanoliter scale. By definition one liter comprises 10^9 nanoliters. Because the claims do not define or limit the dimension of the channel and because any dimension and/or can be defined on the micro scale, the channel of Biosseau et al is encompassed by the claimed channel.

Applicant further argues that the Hepes of Biosseau et al does not react with or bind to sperm as claimed. The argument has been considered but is not found persuasive because as Applicant notes, the Hepes "serves as the media into which the sperm is suspended". Hence, the Hepes reacts with the sperm via suspending the sperm in the media and therefore is encompassed by the claim.

Art Unit: 1634

11. Claims 1-2, 162 and 165 are rejected under 35 U.S.C. 102(b) as being anticipated by Pinkel et al (U.S. Patent No. 5,982,534, issued 9 November 1999).

Regarding Claim 1, Pinkel et al disclose an apparatus for detecting an analyte comprising a suspended beam comprising a microfluidic channel (specimen slide with grooves, Column 5, lines 27-34 **or** channel formed between slide and coverslip, Column 5, lines 37-47) and connected to at least one mechanically stable support (specimen holder, Column 5, lines 48-57) wherein the channel has at least one chemical species (e.g. Column 5, lines 13-26) that reacts with the analyte (sample, Column 8, line 65-column 9, line 13).

Regarding Claim 2, Pinkel et al disclose the apparatus of Claim 1 wherein an analyte binds (Column 5, lines 16-26). The instant claim recites "the microfluidic channel is treated with a capture ligand that binds the analyte". This is a recitation of intended use for the apparatus but does not define or further limit the structure of the apparatus.

Regarding Claim 162, Pinkel et al disclose the apparatus wherein the beam is in a controlled environment e.g. arms for positioning slide (Column 11, lines 19-24).

Regarding Claim 165, Pinkel et al disclose the apparatus wherein the beam is suspended between two mechanically stable supports (#127).

Response to Arguments

12. Applicant argues that Pinkel et al do not teach a microfluidic channel because the dimension of their channel is "at least three orders of magnitude" than that claimed. The argument has been considered but is not found persuasive because the claimed channels are not limited or defined by specific dimensions.

Applicant further argues that Pinkel does not "treat" the channel surface so as to bind the analyte. The argument has been considered but is not found persuasive because Pinkel specifically teaches that arrays of biological polymers e.g. DNA are attached to the viewing surface (Column 5, lines 16-24). However, as stated above, the claim is reasonably interpreted

Art Unit: 1634

as an intended use for the channel of Claim 1. The claim merely requires that the channel be "treated". The claim does not define the channel has having a capture ligand attached to the surface. The claim merely recites a method of treating the channel at some, undefined, time e.g. in the clinical laboratory.

13. Claims 1-7, 19, 28, 43-44 and 161-162 are rejected under 35 U.S.C. 102(e) as being anticipated by Manalis (U.S. Patent Application Publication No. 2004/0038426, filed 2 January 2003).

The applied reference has a common inventor with the instant application. Based upon the earlier effective U.S. filing date of the reference, it constitutes prior art under 35 U.S.C. 102(e). This rejection under 35 U.S.C. 102(e) might be overcome either by a showing under 37 CFR 1.132 that any invention disclosed but not claimed in the reference was derived from the inventor of this application and is thus not the invention "by another," or by an appropriate showing under 37 CFR 1.131.

Regarding Claim 1, Manalis discloses an apparatus comprising at least one beam (cantilever, 115, Fig. 1) suspended and connected to at least one mechanically stable support (support, 120, Fig. 1) wherein the beam contains microfluidic channels (110, Fig. 1) having chemical species for analyte reaction (§ 34-35).

Regarding Claim 2, Manalis discloses the chemical species is a capture ligand (§ 34).

Regarding Claim 3, Manalis discloses the apparatus wherein the beams are resonating and the detector measures resonance (§ 36).

Regarding Claim 4, Manalis disclose the apparatus wherein the capture ligand is bound to the interior surface (§ 33-34).

Regarding Claim 5, Manalis discloses the apparatus further comprising a gel and the ligand is bound to the gel (§ 33).

Art Unit: 1634

Regarding Claim 6, Manalis discloses the apparatus wherein the beam has two channels that meet, then separate (Fig. 1).

Regarding Claim 7, Manalis discloses the apparatus wherein the analyte is transported via pressure from fluid flow (§ 36). While Manalis discloses this functionality, the recited function does not further limit the structural components of the apparatus.

Regarding Claim 19, Manalis discloses the detectors are ligand is a nucleic acid (§ 35)

Regarding Claim 28, Manalis discloses the detector measures conductivity (§ 8).

Regarding Claim 43, Manalis discloses the apparatus wherein the channels have a depth of between 100 and 3000nm (§ 57-59).

Regarding Claim 44, Manalis discloses the apparatus wherein the channels have a thickness of between 100 and 1200nm (§ 57-59).

Regarding Claim 161, Manalis discloses the apparatus further comprising detectors for measuring a change (§ 33 and 39).

Regarding Claim 162, Manalis discloses the apparatus within a controlled environment i.e. multiple beams attached to a single base for calorimetric and mass measurements (§ 26 and Fig. 4).

Response to Arguments

14. Applicant states that a petition has been filed to claim priority to the Manalis reference thereby removing the reference as prior art. As stated above, the petition has been dismissed. The rejection is maintained.

Furthermore, the Manalis reference has a different inventive entity for that of the instant application. Therefore, the Manalis reference is “by another” and would still constitute prior art under 35 U.S.C. 102(e) even if a supplemental petition was submitted and approved.

Art Unit: 1634

15. Claims 1-7, 9-14, 43-44 and 163-164 are rejected under 35 U.S.C. 102(e) as being anticipated by Geli (U.S. Patent Application Publication No. 2003/0027354, filed 10 June 2002).

Regarding Claim 1, Geli discloses an apparatus comprising at least one beam (support #22, perpendicular to fractionation column #20, ¶ 294-298 and Fig. 7) wherein the beam encloses sealed microfluidic channels (#10, ¶ 244 & 270) the microchannels having chemical species for analyte reaction i.e. functional groups and capture molecules (¶ 270-272). Geli further teach measurement of cantilever flex and/or resonance (¶ 262 and 263). The claimed “suspended” is interpreted to encompass the parallel support #22 because the support is illustrated in Fig. 7 as suspending out from the support #20 and is described as “laid out at each side” (¶ 296).

Regarding Claim 161, Geli discloses the apparatus further comprising detectors for measuring a change (¶ 21).

Regarding Claim 2, Geli discloses the chemical species is a capture ligand (¶ 21-23).

Regarding Claim 3, Geli discloses the apparatus wherein the beams are resonating and the detector measures resonance (¶ 263 & 266-272).

Regarding Claim 4, Geli discloses the apparatus wherein the capture ligand is bound to the interior surface or the microchannel associated with the cantilever (¶ 10 and 36).

Regarding Claim 5, Geli discloses the apparatus further comprising a gel and the ligand is bound to the gel (¶ 10).

Regarding Claim 6, Geli discloses the apparatus wherein the beam has two channels that meet, then separate downstream (¶ 10).

Regarding Claim 7, Geli discloses the apparatus wherein the analyte is transported via pressure from fluid flow (¶ 191 and 327). While Geli discloses this functionality, the recited function does not further limit the structural components of the apparatus.

Art Unit: 1634

Regarding Claim 9, Geli discloses the apparatus wherein the resonance is driven by electrodes (¶ 268).

Regarding Claim 10 Geli discloses the apparatus wherein one of the electrodes is common (¶ 268, lines 4-6).

Regarding Claims 11-12 Geli discloses the electrodes are gold via electrical connection to gold cantilevers (¶ 266-268).

Regarding Claim 13, Geli discloses the apparatus wherein the common electrode is in contact with each cantilever (¶ 268, lines 4-8 and ¶ 2690).

Regarding Claim 14, Geli discloses the apparatus wherein the solution is electrolyte (¶ 99, lines 22-26 and ¶ 175).

Regarding Claim 43, Geli discloses the apparatus wherein the channels have a depth of between 100 and 3000nm (¶ 61).

Regarding Claim 44, Geli discloses the apparatus wherein the channels have a thickness of between 100 and 1200nm (¶ 61).

Regarding Claim 163, Geli discloses the apparatus wherein the beam is resonating and the device measures changes in resonance frequency (¶ 263).

Regarding Claim 164, Geli discloses the apparatus wherein the beam is a cantilever (Abstract and ¶ 263).

Response to Arguments

15. Applicant asserts that Geli does not teach a microfluidic channel within the cantilever beam as claimed. The argument has been considered but is not found persuasive because the Geli teaches a beam comprising support #22 projected at a parallel angle from support #20, the support #22 further comprises enclosed channels #10 as claimed.

Art Unit: 1634

Claim Rejections - 35 USC § 103

16. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

17. Claims 15-17, 19-21, 28, 46 and 162-165 are rejected under 35 U.S.C. 103(a) as being unpatentable over Geli (U.S. Patent Application Publication No. 2003/0027354, filed 10 June 2002) in view of Kley (U.S. Patent No. 6,337,479, filed 21 July 1999).

Regarding Claims 46 and 162-165, Geli teaches the apparatus comprising at least one beam (support #22, perpendicular to fractionation column #20, ¶ 294-298 and Fig. 7) wherein the beam encloses sealed microfluidic channels (#10, ¶ 244 & 270) the microchannels having chemical species for analyte reaction i.e. functional groups and capture molecules (¶ 270-272). Geli further teach measurement of cantilever flex and/or resonance (¶ 262 and 263). The claimed "suspended" is interpreted to encompass the parallel support #22 because the support is illustrated in Fig. 7 as suspending out from the support #20 and is described as "laid out at each side" (¶ 296). Geli further teaches the apparatus comprising a cantilever that is resonating whereby the device measures changes in resonance frequency (¶ 263).

Geli teaches the apparatus comprises detection means are selected according to the nature of the desired detection (¶ 21) but they do not teach capacitors or detection of conductivity. Geli is silent regarding a controlled and low-pressure environment for the apparatus.

However, Kley teaches a similar apparatus comprising a suspended cantilever for analyte detection or measurement (Fig. 1-3) wherein the apparatus functions more accurately in a controlled and low-pressure (vacuum) environment (Column 30, lines 35-42). It would have been obvious to one of ordinary skill in the art at the time the claimed invention was

Art Unit: 1634

made to apply the controlled, low-pressure environment of Kley to the apparatus of Geli for the expected benefit of increasing apparatus accuracy as taught by Kley (Column 30, lines 35-42).

Regarding Claim 15-17, Geli teaches the apparatus comprises detection means are selected according to the nature of the desired detection (§ 21) but they do not teach capacitors. Kley teaches their similar apparatus wherein the detectors are capacitors comprising two capacitor plates whereby a gap formed between the plates is measured via a change in voltage between the plates to provides a calibrating measurement (Column 34, lines 1-27). It would have been obvious to one of ordinary skill in the art at the time the claimed invention was made to apply the capacitors of Kley to the apparatus of Geli for the expected benefit of calibrating measurement as taught by Kley (Column 34, lines 1-27).

Regarding Claim 19-21, Geli teaches the apparatus is used for analysis of biological samples (Abstract) but they are silent regarding nucleic acids and single or double stranded DNA. Kley teaches their similar apparatus is used for biological molecules e.g. DNA (Column 9, lines 50-55). It would have been obvious to one of ordinary skill in the art at the time the claimed invention was made to provide nucleic acids (e.g. single or double stranded DNA) to the apparatus of Geli based on their suggested biological applications (Abstract) and further based on the well know use of DNA in the similar apparatus of Kley (Column 9, lines 50-55).

Regarding Claim 28, Geli teaches the cantilever has a conductive coating (e.g. gold) but they do not teach a detector for conductivity. Kley teaches their similar apparatus wherein the cantilever has a conductive coating and wherein the detector measures conductivity of the cantilever to thereby calibrate the apparatus (Column 20, lines 15-45). It would have been obvious to one of ordinary skill in the art at the time the claimed invention was made to modify the apparatus of Geli with the conductivity detector of Kley for the expected benefit of calibrating the apparatus as taught by Kley (Column 20, lines 15-45).

Art Unit: 1634

Response to Arguments

18. Applicant argues that Kley dose not overcome the deficiencies of Geli and therefore the cited art does not teach the claimed invention. The argument has been considered but is not found persuasive for the reasons stated above regarding Geli.

Conclusion

19. No claim is allowed.

20. Any inquiry concerning this communication or earlier communications from the examiner should be directed to BJ Forman whose telephone number is (571) 272-0741. The examiner can normally be reached on 6:00 TO 3:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Ram Shukla can be reached on (571) 272-0735. The fax phone number for the organization where this application or proceeding is assigned is (571) 273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).


Any inquiry of a general nature or relating to the status of this application or proceeding should be directed to (571) 272-0547.

Patent applicants with problems or questions regarding electronic images that can be viewed in the Patent Application Information Retrieval system (PAIR) can now contact the USPTO's Patent Electronic Business Center (Patent EBC) for assistance. Representatives are available to answer your questions daily from 6 am to midnight (EST). The toll free number is (866) 217-9197. When calling please have your application serial or patent number, the type of document you are having an image problem with, the number of pages and the specific nature of the problem. The Patent Electronic Business Center will notify applicants of the resolution of the problem within 5-7 business days. Applicants can also check PAIR to confirm that the problem has been corrected. The USPTO's Patent Electronic Business Center is a complete service center supporting all patent business on the Internet. The USPTO's PAIR system provides Internet-based access to patent application status and history information. It also enables

Art Unit: 1634

applicants to view the scanned images of their own application file folder(s) as well as general patent information available to the public.

For all other customer support, please call the USPTO Call Center (UCC) at 800-786-9199.



BJ Forman, Ph.D.
Primary Examiner
Art Unit: 1634
September 22, 2006